

مینڈرنٹس

۱۔ سی۔ ایم ایچ کوہاٹ کو برائے مالی سال 2021-22 الیکٹروسیڈیکل آلات کی لوکل پرچیز کے سر بمبر کو ٹینڈرز (Quotations) جو کہ انکم ٹیکس اور سیلز ٹیکس ڈیپارٹمنٹ سے رجسٹرڈ ہوں اور جن کا نام FBR کی حالیہ فہرست برائے ٹیکس دہندگان میں شامل ہوں سے مطلوب ہیں۔

۲۔ مینڈر دو مراحل میں کھولا جائے گا، بڈز / کوٹیشنز جمع کروانے کی آخری تاریخ مورخہ 15 اپریل 2022 دن 10 بجے بمقام سی ایم ایچ کوہاٹ ہے۔ تکنیکی پیشکش (Technical Offer) مورخہ 15 اپریل 2022ء کو دن 11 بجے بمقام سی ایم ایچ کوہاٹ اور مالی پیشکش (فائنیشنل آفر بشمول GST 17%) مورخہ 19 اپریل 2022ء کو دن 11 بجے بمقام سی۔ ایم۔ ایچ کوہاٹ میں کھولی جائیں گی۔

۳۔ مینڈر کو سر بمبر لفافوں میں جمع کروایا جائے جس پر واضح طور پر "مینڈر برائے لوکل پرچیز الیکٹروسیڈیکل آلات" درج ہو۔ جس کے اندر مزید دو الگ الگ لفافے تکنیکی پیشکش (Technical Offer) اور مالی پیشکش (Financial Offer) کے نام سے موجود ہونے چاہئیں۔

۴۔ محض تکنیکی لحاظ سے ٹیکنیکل کمیٹی کی منظور شدہ بڈرز (bidders) کی فائنیشنل آفر کھولی جائیں گی۔ تکنیکی کمیٹی کسی بھی بڈ کوٹیشنر (bid / quotation) کو تکنیکی بنیاد پر مسترد کر سکتی ہے اور مسترد کرنے کو وجہ بڈر کو واضح طور پر بتائی جائے گی۔

۵۔ شرائط و ضوابط، مینڈر اپلیکیشن فارم، آلات کی فہرست بمع تکنیکی تفصیلات (Technical Specifications) کسی بھی یوم کار میں دفتری اوقات میں میڈیکل آفسر انچارج کے دفتر سے بعوض مبلغ -/3000 روپے بشکل بینک ڈرافٹ (ناقابل واپسی) بنام کمانڈنٹ کمبائنڈ ملٹری ہسپتال کوہاٹ ادا کر کے حاصل کی جاسکتی ہیں۔

۶۔ بڈرز اپنی فائنیشنل آفر کی کل قیمت کی پانچ فیصد رقم بطور بڈ سیکورٹی (BID-Security/CDR) (قابل واپسی) بشکل بینک ڈرافٹ بنام کمانڈنٹ کمبائنڈ ملٹری ہسپتال کوہاٹ بمع فائنیشنل آفر جمع کروائیں۔

۷۔ مینڈر اشتہارات، آلات کی فہرست اور تکنیکی تفصیلات (Technical Specifications) PPRA کی ویب سائٹ سے بھی حاصل کی جاسکتی ہیں۔

کمانڈنٹ سی ایم ایچ کوہاٹ کینٹ

COMBINED MILITARY HOSPITAL (CMH) KOHAT

TENDER NOTICE

1. Sealed quotations are invited from **authorized EM Eqpt distributors who are registered with income tax & sales tax department and also on active tax payer (ATL) of the Federal Board of Revenue (FBR)**, for tender of local purchase of Electro Medical Equipment for the CFY 2021-22.
2. The tender will be opened in two phases, the last date for submission of bids is 15th April 2022, Uptil 1000 hrs at CMH Kohat. **Opening of technical offers on 15th April 2022 at 1100 hrs at CMH Kohat followed by opening of financial offers (Incl 17%GST) tentatively on 19 April 2022 at 1100 hrs at CMH Kohat.**
3. The bids should be submitted at CMH Kohat in sealed envelopes labeled as **"Quotation for LP of EM EQPT"**, which must contain two separately sealed envelopes labeled as **"Technical Offer"** and **"Financial Offer"**.
4. Financial offer of only the **"Technically Accepted"** bidders shall be opened, duly scrutinized by the **"Technical Committee"** of this hosp. The committee reserves the right to reject a bid on technical grounds after informing the bidder about the reason of rejection.
5. The copy of list of EM Eqpt, tech specifications, terms & conditions and tender application forms can be obtained from office of MO I/C Medical Store on any working day during office hours, on submission of Rs 3000/- (tender documents fee) in the form of bank draft in the name of Commandant CMH Kohat (Non-Refundable).
6. Bidders must also submit bid security / CDR which must be 5% of the quoted amount and be submitted as bank draft (duly verified by bank, Refundable) in the name of Commandant CMH Kohat, also attached with the financial proposal.
7. Tender Notice alongwith list of eqpt and Technical specifications will also be uploaded on PPRA website.

Commandant,
Combined Military Hospital Kohat

CRITICALLY DEFICIENT ITEMS OF EM EQPT – CMH KOHAT

Ser	Name of Equipment	Qty	Specifications	Accessories Required for smooth Function
<u>Dermatology</u>				
1	Liquid Nitrogen Container 30 ltr	01	MAIN FEATURES -High-Strength Aluminum Alloy Structure-Light Empty Weight -Multi-Layer Thermal Isolation Design -Hinged Handle Portable -Optional Locking Cover To Protect Store Biological Samples -Protective Jacket To Prevent Collision Portable -Two-Year Vacuum Warranty	
<u>PEADS DEPT</u>				
	Air Anti Pressure Matters	05	-3 way alternating air cell (ABC) - Cover and cell fabric material nylon TPU with mattress base Nylon / PVC fabric - Elastic Bands for mattress cover	Compressor with air machine
<u>Rehab Dept</u>				
1	CPM Knee	01	Input 100-230 volts AC Treatment time continuous Speed Minimum (100%) = 70 sec \pm 10% (-10°-20°) 75 sec \pm 10%(120°-10°) Mean 60%) =120s \pm 10%(-10°-120°) 120s \pm 10% (120°-10°) Maximum (30%) = 240 sec \pm 10% (-10°-120°) 250 sec \pm 10% (120°-10°) ROM -10° to 120°	1 software CD 2 USB cable 3 Program maker software + user manual 4 machine user manual 5 2x power cable 6 2x connectivity cable 7 Thigh + leg support 8 Foot +ankle support 9 leg + feet Support bands
<u>MED Dept</u>				
1	Potable Suction Machine	02	Double jar Auto cleavable / unbreakable 60 ltr /min suction oil less piston (suction machine)	
2	NIBP Monitor	10	Colour touch screen SpO2,PULSE, NIBP Short / long trend graphic / trend table display automatic brightness adjustment motion & low perfusion automatic SpO2 performance during 8 hrs continuous working on batteries rechargeable	<ul style="list-style-type: none"> • Pitch tone variation for pulse rate • Audio visual alarm adjustable alarm limit
3	Cardiac Monitor	20	I. 12.1" high resolution color TFT LCD display critical parameter II. ECG SPO2, NIBP, Temp, Resp monitored III. 12 leads ECG wave from display and total 11 waves forms display IV. ST analysis, arrhythmia analysis V. Digital SPO2 even for low perfusion cases VI. 72 hours tabular and graphical trends & data storage VII. Durgs dose calculation software VIII. OxyCRG software IX. Upgradable to 12 lead, dual IBP, thermal recorder, ETCO2 X. Multigas (CO2, O2, N2O and 5 anaesthetic agent) & second temperature (rectal) XI. Nellcor technology for pulse oximetry XII. Video output VGA XIII. Network with central monitoring system XIV. Certification : CE/FDA	

OT				
1	Baby Resuscitation Trolley	03	<p>Baby warming System Microprocessor control heating system Mobile with antistatic caster, lockable, bumper guard Manual heat output control 0% to 100% Skin and manual temp control settings Display Range of temp LED Heating Power / source 560 quartz Selection for operating modes skin or manual Pivot arms technology for heating. Head can be moved in both direction allowing X-Ray procedure without moving the baby Integrated Observation lamp Integrated baby bed 680x480 mm approx with secured plexiglass side panels Foldable down, with grid for X-Ray Manual bed inclination Audio and visual alarms for power failure, skin temp, deviations, high temp, skin probe defective / unplugged Lockable anti static caster Left and right cylinder support with suction venture system</p>	<p>-IV Pole -Skin probe (Reusable) -Mattress -Humidifier bott -Flow meter -Hose pipe for air controlling -Hose pipe for O₂ controlling</p>
2	Instrument Trolley	05	<p>1. Stainless steel non magnetic frame 2. two elegant sleeves stainless steel 20 Gauge. 3. pipe hollow stainless steel non magnetic 16 Gauge 4. 4 wheels 3"</p> <p>Dimensions</p> <p>Length 30" Width 20" Height 34" Argon welding</p>	
Surg Dept				
1	Surgical Diathermy	03	Electrosurgical cutting and coagulation bipolar and vessel sealing in single generator	Disposable hand sets of various type
2	Thompson Reactor (Abdominal)	01	Self retaining laparotomy retractor fixed with multiple attachment	-
3	Telescope 10mm 30°	01	<p>Telescope 10mm Camera Head Full HD Image Processing Unit Full HD Fiber Optic Cable Light Source</p>	<p>Verses Needle 10mm Trocars x2 suction irrigation intilation cuff for irrigation cleaning /disinfestations solution</p>
Gyn Dept				
1	Harmonic Focus shear + Adopter tissue technology length 9 cm Ref HAR9F-Ethicon	01	Harmonic focus share + adopter issue technology length 9 cm ref HAR9F-ethicon) required for Leprosopic Genealogical surgery = different difficult TAH	Harmonic Blue Hand Peace Blue Test Tip Ref HPBLUE -1 ETHICON)
2	Harmonic Blue Hand Piece & Blue Test Tip Ref HPBLUE1, Ethicon	01	Harmonic focus share + adopter issue technology length 9 cm ref HAR9F-ethicon) required for Leprosopic Genealogical surgery = different difficult TAH	Harmonic Blue Hand Peace Blue Test Tip Ref HPBLUE -1 ETHICON)

LAB Deptt

Centrifuge 28 holes	01	Rmp 500 6000/min Rcf-4226 Capacity-40x sample tube Made in Germany /Europe Swinging rotor One hand lid lock Lid dropping protection Automatic rotor recognition Error display Running time 1-99min Stainless steel chamber	Stabilizer 3500 watts attached
------------------------	----	--	--------------------------------